

Scope of Accreditation for Calibration

Accreditation No. : CALIBRATION 0156

Laboratory Status : Permanent Site Temporary Mobile

Field of Measurement	Parameter/Range/Item	Best Measurement Capability*	Standard/Technique/Method/Equipment/Remark
1. Electrical (cont.)	Measuring instrument		In-house method : CM-046 by simulate signal to UUC from process calibrator or multi-product calibrator (Electronic part only, not include effect of mechanical part)
	Digital tachometer		
	Photo & contact type (cont.)		
	> 54 000 r/min to 60 000 r/min	1.2 r/min	
	> 60 000 r/min to 66 000 r/min	1.3 r/min	
	> 66 000 r/min to 72 000 r/min	1.5 r/min	
	> 72 000 r/min to 78 000 r/min	1.6 r/min	
	> 78 000 r/min to 84 000 r/min	1.7 r/min	
2. Mass	Conventional mass		In-house method : CM-013 based on OIML R 111-1 : 2004 and UKAS LAB 14 : 2006
	Class F1		
	1 mg	6.0 µg	
	2 mg	6.0 µg	
	5 mg	6.0 µg	
	10 mg	7.0 µg	
	20 mg	7.0 µg	
	50 mg	10 µg	
	100 mg	10 µg	
	200 mg	10 µg	
500 mg	10 µg		
1 g	20 µg		

* expressed as an uncertainty (+) which for k = 2, providing a level of confidence of approximately 95%

Signature

Scope of Accreditation for Calibration

Accreditation No. : CALIBRATION 0156

Laboratory Status : Permanent Site Temporary Mobile

Field of Measurement	Parameter/Range/Item	Best Measurement Capability*	Standard/Technique/Method/Equipment/Remark
2. Mass (cont.)	Conventional mass		In-house method : CM-013 based on OIML R 111-1 : 2004 and UKAS LAB 14 : 2006
	Class F1 (cont.)		
	2 g	20 µg	
	5 g	30 µg	
	10 g	40 µg	
	20 g	40 µg	
	50 g	0.10 mg	
	100 g	0.10 mg	
	200 g	0.20 mg	
	500 g	0.50 mg	
	1 kg	1.0 mg	
	2 kg	2.0 mg	
	5 kg	5.0 mg	
	Class F2		
	10 kg	28 mg	
	20 kg	48 mg	
	Mass		In-house method : CM-013 based on OIML R 111-1 : 2004 and UKAS LAB 14 : 2006
	1 mg	6.0 µg	
	2 mg	6.0 µg	
	5 mg	6.0 µg	
10 mg	7.0 µg		
20 mg	7.0 µg		

* expressed as an uncertainty (+) which for k = 2, providing a level of confidence of approximately 95%

Scope of Accreditation for Calibration

Accreditation No. : CALIBRATION 0156

Laboratory Status : Permanent Site Temporary Mobile

Field of Measurement	Parameter/Range/Item	Best Measurement Capability*	Standard/Technique/Method/Equipment/Remark
2. Mass (cont.)	Mass (cont.)		In-house method : CM-013 based on OIML R 111-1 : 2004 and UKAS LAB 14 : 2006
	50 mg	10 µg	
	100 mg	10 µg	
	200 mg	10 µg	
	500 mg	10 µg	
	1 g	20 µg	
	2 g	20 µg	
	5 g	30 µg	
	10 g	40 µg	
	20 g	40 µg	
	50 g	0.10 mg	
	100 g	0.10 mg	
	200 g	0.20 mg	
	500 g	0.50 mg	
	1 kg	1.0 mg	
	2 kg	2.0 mg	
	5 kg	5.0 mg	
10 kg	28 mg		
20 kg	48 mg		

* expressed as an uncertainty (+) which for k = 2, providing a level of confidence of approximately 95%

Scope of Accreditation for Calibration

Accreditation No. : CALIBRATION 0156

Laboratory Status : Permanent Site Temporary Mobile

Field of Measurement	Parameter/Range/Item	Best Measurement Capability*	Standard/Technique/Method/Equipment/Remark
2. Mass (cont.)	Electronic balance		In-house method : CM-014 based on UKAS LAB 14 : 2006
	Up to 10 mg	20 µg	
	> 10 mg to 20 mg	22 µg	
	> 20 mg to 50 mg	24 µg	
	> 50 mg to 100 mg	29 µg	
	> 100 mg to 200 mg	34 µg	
	> 200 mg to 500 mg	41 µg	
	> 500 mg to 1 g	49 µg	
	> 1 g to 2 g	63 µg	
	> 2 g to 5 g	78 µg	
	> 5 g to 10 g	94 µg	
	>10 g to 20 g	0.13 mg	
	> 20 g to 50 g	0.17 mg	
	> 50 g to 100 g	0.28 mg	
	> 100 g to 200 g	0.52 mg	
	> 200 g to 500 g	1.4 mg	
	> 500 g to 1 kg	3.1 mg	
	> 1 kg to 2 kg	5.4 mg	
> 2 kg to 5 kg	14 mg		
> 5 kg to 10 kg	31 mg		
> 10 kg to 20 kg	54 mg		
> 20 kg to 50 kg	0.14 g		

* expressed as an uncertainty (+) which for k = 2, providing a level of confidence of approximately 95%

Handwritten signature

Scope of Accreditation for Calibration

Accreditation No. : CALIBRATION 0156

Laboratory Status : Permanent Site Temporary Mobile

Field of Measurement	Parameter/Range/Item	Best Measurement Capability*	Standard/Technique/Method/Equipment/Remark
2. Mass (cont.)	Electronic balance (cont.)		In-house method : CM-014 based on UKAS LAB 14 : 2006
	> 50 kg to 100 kg	1.5 g	
	> 100 kg to 200 kg	3.7 g	
	> 200 kg to 500 kg	11 g	In-house method : CM-051 based on UKAS LAB 14 : 2006
	Mechanic balance		
	Up to 500 g	0.90 g	
	> 500 g to 1 kg	1.0 g	
	> 1 kg to 2 kg	2.0 g	
	> 2 kg to 5 kg	4.0 g	
	> 5 kg to 7 kg	5.0 g	
	> 7 kg to 30 kg	12 g	
	> 30 kg to 50 kg	17 g	
	> 50 kg to 100 kg	58 g	
	> 100 kg to 300 kg	58 g	
	> 300 kg to 500 kg	83 g	
3. Temperature	Digital thermometer with sensor Resistance Temperature Detector (RTD)		In-house method : CM-003 by comparison with PRT in liquid bath and dry block
	-30 °C to 50 °C	0.14 °C	
	> 50 °C to 400 °C	0.19 °C	
	> 400 °C to 600 °C	0.21 °C	

* expressed as an uncertainty (+) which for k = 2, providing a level of confidence of approximately 95%

Signature